

SUPERVISOR TEST

STONY POINT
OWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 01/18/2026

OPERATOR NAME:
CHRISTOPHER R JORDAN
Permit #: 46293
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:22
INTERNAL STANDARD	VERIFIED	02:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:23
BLANK TEST	0.00	02:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:24
BLANK TEST	0.00	02:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:25
BLANK TEST	0.00	02:26

